



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/864,891	05/24/2001	Kok-Hwee Ng	F-5735 (1417P P 599)	2261

7590 03/03/2006
Bradford R. L. Price Esq.
Senior Counsel
Baxter International Inc.
Route 120 and Wilson Road, RLP-30
Round Lake, IL 60073

EXAMINER

TOMASZEWSKI, MICHAEL

ART UNIT PAPER NUMBER

3626

DATE MAILED: 03/03/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/864,891

Applicant(s)

NG ET AL.

Examiner

Mike Tomaszewski

Art Unit

3626

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) Responsive to communication(s) filed on 09 September 2005.
- 2a) This action is FINAL. 2b) This action is non-final.
- 3) Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) Claim(s) 8-54 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) Claim(s) _____ is/are allowed.
- 6) Claim(s) 8-54 is/are rejected.
- 7) Claim(s) _____ is/are objected to.
- 8) Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) The specification is objected to by the Examiner.
- 10) The drawing(s) filed on 24 May 2001 is/are: a) accepted or b) objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
a) All b) Some * c) None of:
1. Certified copies of the priority documents have been received.
 2. Certified copies of the priority documents have been received in Application No. _____.
 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| Paper No(s)/Mail Date <u>27 August 2004</u> . | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Notice To Applicant

1. This communication is in response to the application filed on 24 May 2001 and the amendment filed 09 September 2005. Examiner notes that Claim 54 was inadvertently excluded from the first non-final action filed on 07 June 2005. Therefore, claims 8-54 have been re-examined herein and are pending. The IDS statements filed on 27 August 2004, 26 August 2004, 09 March 2004, 13 August 2003, 11 July 2003, 17 January 2003, 04 September 2001, and 29 August 2001 have been entered and considered.

Claim Rejections - 35 USC § 103

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Art Unit: 3626

3. Claims 1-54 are rejected under 35 U.S.C. 103(a) as being unpatentable over Fletcher-Haynes et al. (US 2001/0034614; hereinafter Fletcher), in view of Otworth (et al. (US 2002/0059030; hereinafter Otworth).

(A) As per claim 8, Fletcher discloses a system for managing inventory of blood component collection soft goods and for preventing the use of quarantined soft goods, the system comprising:

- (a) a blood component collection instrument for collecting a blood component from a blood component donor (Fletcher: par. [0056]);
- (b) a system computer being operably connected to the blood component collection instrument (Fletcher: par. [0057]), the system computer running a blood component collection application for at least a portion of a blood component collection process (Fletcher: par. [0057]), wherein the system computer is in data communication with a system database having a blood component collection soft good inventory (Fletcher: par. [0056], [0063], and [0195]) (Examiner considers a blood component (e.g., platelets, red blood cells, stem cells, white blood cells, plasma) to read on "blood component collection soft good."); and,
- (c) an interface being operably connected to the system computer (Fletcher: par. [0057]).

Examiner notes that Fletcher teaches the use of a multitude of graphical user interfaces (GUIs) having numerous fields for a indicating a variety of things including blood component inventory levels, machine identification numbers, blood collection bag identification numbers, blood component collection volumes, tubing set identification numbers, ad nauseum (Fletcher: Figs. 2A-6M). As such, Examiner considers GUIs having various inventory indication fields to be notoriously well known.

Fletcher, however fails to expressly disclose a system for managing inventory of blood component collection soft goods and for preventing the use of quarantined soft goods, the system comprising:

- (d) the interface having a [*quarantine field*] for indicating that at least a portion of the blood component collection soft good inventory is [*quarantined*].

Nevertheless, this feature (i.e, quarantining) is old and well known in the art, as evidenced by Otworth. In particular, Otworth discloses a system for managing inventory of blood component collection soft goods and for preventing the use of quarantined soft goods, the system comprising:

- (d) the interface having a quarantine field for indicating that at least a portion of the blood component collection soft good inventory is quarantined (Otworth: par. [0222]; Fig. 16).

One of ordinary skill would have found it obvious at the time of the invention to combine the teachings of Otworth (e.g., quarantining, in particular) with the teachings of Fletcher (e.g., interfaces, having various inventory fields, operably connected to the system computer) with the motivation of more efficiently monitoring and controlling inventories of products (e.g., soft goods) (Otworth: par. [0028]).

(B) As per claim 9, Fletcher discloses the system of claim 8, wherein the interface communicates to the system database an identification of the soft goods (Fletcher: par. [0022], [0083], [0125], and [0162]) (Examiner notes also that Fletcher teaches the use of various GUI comment fields whereby a user of the system could indicate that a particular soft good is unsuitable (i.e., quarantined)).

Fletcher, however, fails to expressly disclose the system of claim 8, wherein the interface communicates to the system database an identification of the [quarantined] soft goods.

Nevertheless, this feature, as aforementioned, is old and well known, as evidenced by Otworth. In particular, Otworth discloses the system of claim 8, wherein the interface communicates to the system database an identification of the quarantined soft goods (Otworth: par. [0222]; Fig. 16).

One of ordinary skill would have found it obvious at the time of the invention to combine the teachings of Otworth (e.g., quarantining, in particular) with the teachings of Fletcher (e.g., interfaces, having various inventory fields, operably connected to the

Art Unit: 3626

system computer) with the motivation of more efficiently monitoring and controlling inventories of products (e.g., soft goods) (Otworth: par. [0028]).

(C) As per claim 10, Fletcher discloses the system claim 8, wherein the blood component collection soft good is selected from a group consisting of blood component collection kit (Fletcher: par. [0022], [0071], [0192], and [0315]) (Examiner considers a barcode, needle, receptacle bag, tubing set, and blood containers, among other soft goods to read on "blood component collection kit.").

Examiner has noted insofar as claim 10 recites "selected from a group consisting of blood component collection kit, a blood component collection solution, and a blood component collection transfer pack," a blood component collection kit has been recited.

(D) As per claim 11, Fletcher discloses the system of claim 8, wherein the interface further comprises a reader being operably connected to the system computer for receiving an operator identifier and transmitting the operator identifier to the system computer, and for receiving separate input of a blood component soft good identifier and transmitting the blood component soft good identifier to the system database (Fletcher: par. [0022], [0059], [0071], [0079], [0083], and [0125]).

In short, Fletcher teaches a system replete with a myriad of identifiers (e.g., operator identifiers, instrument/device/identifiers, donor identifiers, soft good identifiers, inventory identifiers, etc.) being received and transmitted by the system computer in

Art Unit: 3626

conjunction with an array of peripheral devices (e.g., barcode readers, scanners, cameras, etc.) (Fletcher: par. [0022], [0059], [0071], [0079], [0083], and [0125]).

(E) As per claim 12, Fletcher discloses the system of claim 11, wherein the operator identifier and a blood component collection soft good identifier are received from a location proximate the blood component collection instrument (Fletcher: par. [0022], [0058], [0059], [0079], [0083], and [0125]).

(F) As per claim 13, Fletcher discloses the system claim 8, wherein the system database is integral with the system computer (Fletcher: par. [0057] and [0058]).

In fact, Fletcher teaches multiple system configurations including, but not limited to, one where the system database can be included therein (i.e., integral within the computer), one where the system database is positioned in close proximity with the system computer (i.e., integral to the computer system), and one where the system database is located remotely (i.e., integral to the computer system network) (Fletcher: par. [0057] and [0058]).

(G) As per claim 14, Fletcher discloses the system of claim 8, further comprising a blood component collection donor identifier corresponding to a blood component donor, wherein the blood component collection donor identifier is transmittable to the system computer for storing the blood component collection donor identifier in the memory and for associating the blood component collection donor identifier with at least one of the

Art Unit: 3626

blood component collection soft good identifier and the blood collection instrument identifier (Fletcher: par. [0022], [0068], [0124], [0125], [0142] and [0159]; Fig. 2A-6M).

(H) As per claim 15, Fletcher discloses the system of claim 8, wherein the blood component collection instrument further comprises a blood component collection instrument identifier (Fletcher: par. [0159]; Fig. 4A).

(I) As per claim 16, Fletcher discloses the system claim 8, wherein the interface utilizes radio frequency to transmit to the system computer (Fletcher: par. [0059]).

In fact, Fletcher teaches an open computer system architecture that may leverage a broad assortment of interface transmission means including cable, satellite, and energy wave communication, among other transmission means (Fletcher: par. [0059]).

(J) As per claim 17, Fletcher discloses the system of claim 8, further comprising:

- (a) a system communication conduit for operably connecting the system computer to the blood component collection instrument (Fletcher: par. [0012]); and,

Art Unit: 3626

- (b) a system communication protocol for facilitating communication on the communication conduit between the system computer and the blood component collection instrument (Fletcher: par. [0030]).
- (K) As per claim 18, Fletcher discloses the system of claim 17, wherein the system communication protocol is Ethernet (Fletcher: par. [0030]).
- (L) As per claim 19, Fletcher discloses the system of claim 17, wherein the system communication protocol is TCP/IP (Fletcher: par. [0030] and [0194]).
- (M) As per claim 20, Fletcher discloses the system of claim 17, further comprising:
- (a) a network server being operably connected to the system computer via a network communication conduit (Fletcher: par. [0012] and [0065]); and
 - (b) a web interface being operably connected to the system computer for facilitating access to the blood component collection process, wherein the interface receives data from the system computer (Fletcher: par. [0033] and [0194]).
- (N) As per claim 21, Fletcher discloses the system of claim 20, further comprising a web server being operably connected to the system computer and operably responsive

Art Unit: 3626

to a web browser wherein the information stored in the system computer can be accessed (Fletcher: par. [0033] and [0194]).

(O) As per claim 22, Fletcher discloses the system of claim 20, wherein the interface comprises a reader having at least one of a touch pad (Fletcher: par. [0057]).

Examiner has noted insofar as claim 22 recites "at least one of a touch pad, a keypad, an optical scanner, and a magnetic scanner" a touch pad has been recited.

(P) As per claim 23, Fletcher discloses the system of claim 8, wherein the system database further comprises separate inventory data for each of a plurality of different types of soft goods (Fletcher: par. [0011] and [0166]; Figs. 2A-6M).

Examiner notes that Fletcher specifically teaches capturing, tracking, editing, printing, manipulating, measuring, modifying, calculating, transmitting, and receiving various soft goods (e.g., various blood component solutions, such as plasma, red blood cells, etc.; blood collection kits including needles, blood collection bags, etc.) inventory data pertinent to the blood collection process (Fletcher: par. [0011] and [0166]; Figs. 2A-6M).

(Q) As per claim 24, Fletcher discloses the system of claim 23, wherein the plurality of different types of soft goods is a blood component collection kit (Fletcher: par. [0022], [0071], [0192], and [0315]) (Examiner considers a barcode, needle, receptacle bag,

Art Unit: 3626

tubing set, and blood containers, among other soft goods to read on “blood component collection kit.”).

Examiner has noted insofar as claim 24 recites “selected from group consisting of a blood component collection kit, a blood component collection solution, and a blood component collection transfer pack” a blood component collection kit has been recited.

(R) As per claim 25, Fletcher discloses the system of claim 8, wherein the blood component soft good inventory data is modified in response to the receipt of the blood component soft good identifier transmitted from the interface (Fletcher: par. [0022], [0084], [0125] and [0166]; Figs. 2A-6M).

(S) As per claim 26, Fletcher discloses the system of claim 25, wherein the system computer generates a notification when the blood component soft good inventory data is modified to a value which is lower than a predetermined value (Fletcher: par. [0314]) (Examiner considers blood component collection solutions (e.g., plasma solutions, red blood cell solutions, etc.) to read on “blood component soft good.”

(T) As per claim 27, Fletcher discloses the of claim 26, wherein the notification comprises providing a reorder option corresponding to the blood component soft good associated with the blood component soft good identifier (Fletcher: par. [0195]).

Art Unit: 3626

(U) As per claim 28, Fletcher discloses the system of claim 27, wherein the notification is transmitted to a remote access service for restocking blood component soft good inventory (Fletcher: par. [0195] and [0314]).

(V) As per claim 29, Fletcher discloses the system of claim 8, further comprising a blood component collection kit having a plurality of blood component collection soft goods (Fletcher: par. [0022], [0056], [0063], [0071], [0192], [0195], [0395]) (Examiner considers a blood component (e.g., platelets, red blood cells, stem cells, white blood cells, plasma) to read on "blood component collection soft good.")

(W) As per claim 30, Fletcher discloses the system of claim 29, wherein the blood component collection kit comprises a blood component container, a hypodermic needle, a blood component sample container, and a label (Fletcher: par. [0022], [0071], [0192], and [0315]).

Examiner considers a barcode (i.e., label), needle, receptacle bag (i.e., blood component container, blood component sample container), tubing set, and blood containers (i.e., sample container), among other soft goods to read on "blood component collection kit."

(X) Claim 31 differs from system claim 8 by reciting "[a] computer readable medium having computer program code stored thereon..." within its preamble. As per these elements, Fletcher's system and method for managing inventory of blood component

Art Unit: 3626

collection soft goods includes computers, data storage devices, communication devices, server systems, network systems and software applications running in conjunction with various hardware devices (Fletcher: par. [0020], [0031], [0032] and [0057]). As such, it is readily apparent that Fletcher's system and method for managing the inventory of blood collection soft goods is controlled by a computer program stored upon a computer-readable medium.

The remainder of claim 31 substantially repeats the same limitations of claim 8 and is therefore, rejected for the same reasons given for claim 8 above and incorporated herein.

(Y) Claims 32-42 substantially repeat the same limitations of claims 9-12, 14-15 and 25-28, and are therefore, rejected for the same reasons given for those claims.

(Z) Method claim 43 differs from system claim 8 by excluding hardware and software elements, namely, "a blood component collection instrument," "a system computer being operably connected to the blood collection instrument," "the system computer running a blood component collection application," "a system database having a blood component collection soft good inventory," and "an interface being operably connected to the system computer, the interface having a quarantine field." The method merely repeats the underlying process steps of system claim 8 and thus, merely repeats the same limitations of claims 8 and is therefore, rejected for the same reasons given for claim 8 above and incorporated herein.

(AA) Claims 44-53 substantially repeat the same limitations of claims 9-12, 14-15 and 25-27, and are therefore rejected for the same reasons given for those claims.

Conclusion

4. The prior art made of record and not relied upon is considered pertinent to Applicant's disclosure. The cited but not applied art teaches a tissue specimen collection kit (5,424,040) and an inventory control system (4,961,533).

The cited but not applied prior art also includes non-patent literature articles by Selznick, Sanford H. ("Development and Application of Computer Software For Cell Culture Laboratory Management" Sep. 8, 2000.) **discussing software applications having data fields to flag whether an item is in quarantine**; Stokes, Teri. ("The Survive and Thrive Guide to Computer Validation" 1998. Interpharm/CRC.) **discussing quarantining of inventory**; Ballard, R. L. ("Methods of Inventory Monitoring and Measurement" Feb 20, 1996. The Logistics Business. pg. 1-17.) **discussing inventory quarantining methods and software**; Hellman, Rita ("A Fictitious HIS-Reconstruct of an Inventory Information System" Feb 1989. Computers in Industry. Vol. 11, Iss. 4. pg. 301.) **discussing quarantine inventory management**; The Patriot News ("AIDS: Risky Transfusions Spur Lab-Procedure Change" Jul 31, 1987. pg. B.3.) **discussing**


quarantining unsuitable biological items; and O'Rourke, Kevin ("FDA Approves Home Test To Detect Hepatitis C" Jun 7, 1999. Drug Store News. Vol. 21, Iss. 9. pg. 154.) **discussing blood collection kits.**


5. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Mike Tomaszewski whose telephone number is (571)272-8117. The examiner can normally be reached on M-F 7:00 am - 3:30 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Joseph Thomas can be reached on (571)272-6776. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

MT

 11.18.05


JOSEPH THOMAS
SUPERVISORY PATENT EXAMINER